



K&S Associates, Inc.
 1926 Elm Tree Drive
 Nashville, Tennessee 37210-3718
 800-522-2325 Fax 615-871-0856



DIAGNOSTIC X-RAY CALIBRATION ENERGIES

Facility _____ Phone _____

Person _____ Fax _____

Chamber/System Description _____

Combined Expanded Uncertainties (including NIST) at the 95% confidence level:

Mammography: 3%

Rad/Fluoro/CT: 5%

Survey Meter X-ray point: 10%

Mammography*

- ___ 24 kVp (0.27 mm Al HVL)
- ___ 25 kVp (0.29 mm Al HVL)
- ___ 26 kVp (0.31 mm Al HVL)
- ___ 28 kVp (0.33 mm Al HVL)
- ___ 30 kVp (0.35 mm Al HVL)
- ___ 32 kVp (0.36 mm Al HVL)
- ___ 35 kVp (0.38 mm Al HVL)

Radiographic

- ___ 50 kVp (1.83 mm Al HVL)
- ___ 60 kVp (2.20 mm Al HVL)
- ___ 80 kVp (2.90 mm Al HVL)
- ___ 100 kVp (3.70 mm Al HVL)
- ___ 120 kVp (4.49 mm Al HVL)
- ___ 140 kVp (5.37 mm Al HVL)

Fluoroscopic

- ___ 50 kVp (1.83 mm Al HVL)
- ___ 60 kVp (2.20 mm Al HVL)
- ___ 80 kVp (2.90 mm Al HVL)
- ___ 90 kVp (3.31 mm Al HVL)
- ___ 100 kVp (3.70 mm Al HVL)

CT

- ___ 120 kVp (8.34 mm Al HVL)
- ___ 140 kVp (9.18 mm Al HVL)

Survey Meter**

- ___ 70 kVp (8.0mm Al HVL)
- ___ 90 kVp (9.5 mm Al HVL)
- ___ 110 kVp (10.7 mm Al HVL)

Low Rate Dosimeters

- ___ 70 kVp (8.0 mm Al HVL)
- ___ 90 kVp (9.5 mm Al HVL)
- ___ 110 kVp (10.7 mm Al HVL)

*Complies with MQSA requirements for NIST measurement assurance agreement- 900.2(xx) and for NIST traceability 900.12(d)(12)

** Single point calculation. Survey meter linearity is established with a separate Cs-137 calibration

Calibration intervals: Your re-calibration reminder will be set on a 12 month interval unless you specify otherwise. IF you wish a re-calibration date recorded on your report please indicate the interval you require in the space provided. _____

American Express, MasterCard, Visa are accepted

email: ksinfo@kslab.com

web: <http://www.kslab.com>