



K&S Associates, Inc.
 1926 Elm Tree Drive
 Nashville, Tennessee 37210-3718
 800-522-2325 Fax 615-871-0856



KEITHLEY 35050A/35080 SYSTEM

FACILITY: _____

CONTACT PERSON: _____

FAX: _____ PHONE: _____

SPECIAL Kit Calibration
 includes the following:

precalibration evaluation of all kit components

35080 kVp divider noninvasive kVp meter calibration, includes all filter packs

diagnostic range - W target, Al filter, 50 - 150 kVp, (3.48 mm Al HVL at 80 kVp)

mammography range - Mo target, Mo filter, 22 - 35 kVp (.35mm Al @ 28 kVp)

35050 dosimeter calibration includes entry of calibration factors for each chamber

96035 15cc chamber ADCL calibration

28 kVp Mo/Mo (MQSA compliant)

100 kVp general rad

96020 150cc chamber ADCL calibration

100 kVp general rad

timer accuracy

mA/mAs meter accuracy

Calibration intervals:

Your re-calibration reminder will be set up on a “manufacturer’s recommended” interval unless you specify otherwise. IF you wish a re-calibration date recorded on your report please indicate the interval you require in the space below.

American Express, Visa or MasterCard are accepted.
 Shipping and handling is not included in the above fees.

Phone 800-522-2325 Fax 615-871-0856
 email: ksinfo@kslab.com web: <http://www.kslab.com>