



**K&S Associates, Inc.**  
 1926 Elm Tree Drive  
 Nashville, Tennessee 37210-3718  
 800-522-2325 Fax 615-871-0856



## DIAGNOSTIC X-RAY CALIBRATION ENERGIES

Facility \_\_\_\_\_ Phone \_\_\_\_\_

Person \_\_\_\_\_ E-Mail \_\_\_\_\_

Chamber/System Description \_\_\_\_\_

Combined Expanded Uncertainties (including NIST) at the 95% confidence level:

Mammography: 3%

Rad/Fluoro/CT: 5%

Survey Meter X-ray point: 10%

**Mammography\***

- \_\_\_ 24 kVp (0.27 mm Al HVL)
- \_\_\_ 25 kVp (0.29 mm Al HVL)
- \_\_\_ 26 kVp (0.31 mm Al HVL)
- \_\_\_ 28 kVp (0.33 mm Al HVL)
- \_\_\_ 30 kVp (0.35 mm Al HVL)
- \_\_\_ 32 kVp (0.36 mm Al HVL)
- \_\_\_ 35 kVp (0.38 mm Al HVL)

**Radiographic**

- \_\_\_ 50 kVp (1.83 mm Al HVL)
- \_\_\_ 60 kVp (2.20 mm Al HVL)
- \_\_\_ 80 kVp (2.90 mm Al HVL)
- \_\_\_ 100 kVp (3.70 mm Al HVL)
- \_\_\_ 120 kVp (4.49 mm Al HVL)
- \_\_\_ 140 kVp (5.37 mm Al HVL)

**Fluoroscopic**

- \_\_\_ 50 kVp (1.83 mm Al HVL)
- \_\_\_ 60 kVp (2.20 mm Al HVL)
- \_\_\_ 80 kVp (2.90 mm Al HVL)
- \_\_\_ 90 kVp (3.31 mm Al HVL)
- \_\_\_ 100 kVp (3.70 mm Al HVL)

**CT**

- \_\_\_ 120 kVp (8.34 mm Al HVL)
- \_\_\_ 140 kVp (9.18 mm Al HVL)

**Survey Meter\*\***

- \_\_\_ 70 kVp (8.0mm Al HVL)
- \_\_\_ 90 kVp (9.5 mm Al HVL)
- \_\_\_ 110 kVp (10.7 mm Al HVL)

**Low Rate Dosimeters**

- \_\_\_ 70 kVp (8.0 mm Al HVL)
- \_\_\_ 90 kVp (9.5 mm Al HVL)
- \_\_\_ 110 kVp (10.7 mm Al HVL)

\*Complies with MQSA requirements for NIST measurement assurance agreement- 900.2(xx) and for NIST traceability 900.12(d)(12)

\*\* Single point calculation. Survey meter linearity is established with a separate Cs-137 calibration

**Calibration intervals: Your re-calibration reminder will be set on a 12 month interval unless you specify otherwise. If you wish a re-calibration date recorded on your report please indicate the interval you require in the space provided.** \_\_\_\_\_

*American Express, MasterCard, Visa are accepted*

email: [ksinfo@kslab.com](mailto:ksinfo@kslab.com)

web: <http://www.kslab.com>